

Technical Data Sheet

Pressure • Temperature • Humidity • Air Velocity • Air Flow • Acoustics



Air quality **AQ 100**









Functions

- Selection of units (temperature)
- Minimum and maximum values
- HOLD function
- Adjustable automatic shut-off
- Adjustable backlight

Technical features

Measuring element	CO ₂ : Infrared sensor.	
	(Non dispersive- infrared sensor NDIR)	
	Ambient temperature : Pt100 class A	
Display	2 lines, LCD technology. Sizes 50 x 34.9 mm.	
	1 line of 5 digits with 7 segments (value)	
	1 line of 5 digits with 16 segments (unit)	
Housing	Shock-proof made of ABS, IP54 protection	
Keypad	Metal-coated with 5 keys	
Cable	retractable, length 450 mm, up to 2.4 m when	
	released	
Conformity	electromagnetical compatibility (NF EN 61326-1 guideline)	
Power supply	1 alcaline battery 9V 6LR61	
Operating temperature	from 0 to 50°C	
Storage temperature	from -20 to +80°C	
Auto shut-off	.adjustable from 0 to 120 min	
Weight	190g	
Languages	French, english	




AQ100 - CO / temperature probe - Fixed probe




Working principle


NDIR - Non dispersive infrared absorbance

A gas absorbs light at a specific wavelength, some of the intensity emitted by the infrared source is absorbed by the gas sample. The amount of light read by the IR sensor is inversely proportional to ${\rm CO}_2$ concentration.



Dimensions





Supplied with ...

DESCRIPTION	AQ 100	
Calibration certificate*	•	
Transport case	•	



^{*}except class 102S



Accessories (See related datasheet)

CE 100	RTS
Protective cover with magnet and holding system	Telescopic extension (for probe), 1m long and bent at 90°.

■ Warranty period

Instruments have 1-year guarantee for any manufacturing defect (return to our After-Sales Service required for appraisal).

www.kimo.fr

EXPORT DEPARTMENT

Tel: + 33. 1. 60. 06. 69. 25 - Fax: + 33. 1. 60. 06. 69. 29 e-mail: export@kimo.fr



Distributed by :

^{*}All accuracies indicated in this document were stated in laboratory conditions and can be guaranteed for measurements carried out in the same conditions, or carried out with required compensation.
**the accuracy is expressed either by a deviation in ppm, or by a percentage of the value concerned. Only the bigger value is considered.